

C

Capabilities

Contents

| | |
|--|----|
| Electrical – Voltage | 2 |
| DC Voltage Source | 2 |
| DC Voltage Measure | 2 |
| AC Voltage Source | 2 |
| AC Voltage Measure | 3 |
| Electrical Simulation of Thermocouple | 5 |
| Electrical – Current | 6 |
| DC Current - Source | 6 |
| DC Current - Measure | 6 |
| AC Current - Source | 7 |
| AC Current - Measure | 7 |
| Electrical – Resistance | 8 |
| Resistance - Source | 8 |
| Resistance Measure | 9 |
| Time and Frequency – Frequency / Period | 10 |
| Oscilloscope | 11 |
| Electrical-RF / Microwave | 12 |
| Generators | 12 |
| Pulse Measurement | 14 |
| Spectrum Analyzers | 15 |
| Scattering Parameters | 16 |
| Length – Artifacts and Standards 1D | 17 |
| Length – Hand Tools and Precision Gages 1D | 17 |
| Mass – Pressure/Low Vacuum | 18 |
| Mass – Scale and Balances | 18 |
| Mass – Torque | 18 |

Electrical - Voltage

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|-------------------------------------|------------------|---|-----------------------------|
| DC Voltage Source | (0 to 200) mV | 5.67 µV/V + 1.0 µV | Wavetek 4808 Opt 10 / 30 |
| | (0.1 to 2.0) V | 5.5 µV/V + 1.4 µV | |
| | (1 to 20)V | 4.4 µV/V + 2.7 µV | |
| | (20 to 200) V | 5.6 µV/V + 64.7 µV | |
| | (100 to 1100) V | 7.36 µV/V + 591 µV | |
| DC Voltage Measure | (0 to 120) mV | 9.4 µV/V + 1.0 µV | Agilent 3458A |
| | (0.1 to 1.2) V | 8.0 µV/V + 0.86 µV | |
| | (1 to 12) V | 8.2 µV/V + 2.7 µV | |
| | (10 to 120)V | 10.5 µV/V + 59 µV | |
| | (100 to 1050) V | 12.2 µV/V + 168 µV | |
| | (1 to 6) kV | 11.6 mV/V + 40 mV | |
| AC Voltage Source (1 to 20) mV | (10 to 31) Hz | 169 µV/V + 7.8 µV | Fluke 80k-6 / 87V |
| | (32 to 330) Hz | 57.4 µV/V + 9.0 µV | |
| | (0.3 to 10) kHz | 129 µV/V + 6.7 µV | |
| | (10 to 33) kHz | 127 µV/V + 6.7 µV | |
| | (30 to 100) kHz | 266 µV/V + 8.4 µV | |
| | (100 to 330) kHz | 1.3 mV/V + 13.3 µV | |
| | (0.3 to 1) MHz | 2.5 mV/V + 29.3 µV | |
| AC Voltage Source (10 to 200) mV | (10 to 31) Hz | 139 µV/V + 14.1 µV | Wavetek 4808 Opt 20 / 30 |
| | (32 to 330) Hz | 27.4 µV/V + 14.7 µV | |
| | (0.3 to 10) kHz | 81.6 µV/V + 11.6 µV | |
| | (10 to 33) kHz | 81.6 µV/V + 11.6 µV | |
| | (30 to 100) kHz | 324 µV/V + 19.9 µV | |
| | (100 to 330) kHz | 1.1 mV/V + 30.2 µV | |
| | (0.3 to 1) MHz | 2.4 mV/V + 157 µV | |
| AC Voltage Source (0.1 to 2) V | (10 to 31) Hz | 99 µV/V + 70 µV | |
| | (32 to 330) Hz | 62.4 µV/V + 71.2 µV | |
| | (0.3 to 33) kHz | 54 µV/V + 45 µV | |
| | (30 to 100) kHz | 90.8 µV/V + 74.4 µV | |
| | (100 to 330) kHz | 260 µV/V + 146 µV | |
| | (0.3 to 1) MHz | 1.8 mV/V + 922 µV | |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

Electrical – Voltage

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|--|------------------|---|-----------------------------|
| AC Voltage Source (1 to 20) V | (10 to 31) Hz | 99.1 μ V/V + 708 μ V | Wavetek 4808 Opt 20 / 30 |
| | (32 to 330) Hz | 62.5 μ V/V + 721 μ V | |
| | (0.3 to 33) kHz | 44 μ V/V + 350 μ V | |
| | (30 to 100) kHz | 91.5 μ V/V + 829 μ V | |
| | (100 to 330) kHz | 262 μ V/V + 1.8 mV | |
| | (0.3 to 1) MHz | 1.7 mV/V + 1.2 mV | |
| AC Voltage Source (10 to 200) V | (10 to 31) Hz | 130 μ V/V + 7 mV | Wavetek 4808 Opt 20 / 30 |
| | (32 to 330) Hz | 70.4 μ V/V + 6.8 mV | |
| | (0.3 to 10) kHz | 53.9 μ V/V + 3.1 mV | |
| | (10 to 33) kHz | 69.3 μ V/V + 3.6 mV | |
| | (30 to 100) kHz | 131 μ V/V + 9.6 mV | |
| | (100 to 330) kHz | 843 μ V/V + 34.5 mV | |
| AC Voltage Source (100 to 1100) V 750V max | (10 to 330) Hz | 141 μ V/V + 26.2 mV | Wavetek 4808 Opt 20 / 30 |
| | (0.3 to 3.3) kHz | 92.9 μ V/V + 35.8 mV | |
| | (3 to 10) kHz | 93.5 μ V/V + 31.2 mV | |
| | (10 to 33) kHz | 131 μ V/V + 34.6 mV | |
| | (30 to 100) kHz | 998 μ V/V + 92.2 mV | |
| | (1 to 40) Hz | 293 μ V/V + 5.7 μ V | |
| AC Voltage Measure (1 to 12) mV | 40 Hz to 1 kHz | 123 μ V/V + 4.2 μ V | Agilent 3458A |
| | (1 to 20) kHz | 335 μ V/V + 3.1 μ V | |
| | (20 to 50) kHz | 888 μ V/V + 3.1 μ V | |
| | (50 to 100) kHz | 4.7 mV/V + 3.5 μ V | |
| | (100 to 300) kHz | 37.8 mV/V + 11.4 μ V | |
| | (1 to 40) Hz | 83.6 μ V/V + 11 μ V | |
| AC Voltage Measure (10 to 120) mV | 40 Hz to 1 kHz | 61.4 μ V/V + 5.3 μ V | Agilent 3458A |
| | (1 to 20) kHz | 121 μ V/V + 5.2 μ V | |
| | (20 to 50) kHz | 275 μ V/V + 5 μ V | |
| | (50 to 100) kHz | 690 μ V/V + 13.9 μ V | |
| | (100 to 300) kHz | 2.8 mV/V + 24.9 μ V | |
| | (0.3 to 1) MHz | 9.1 mV/V + 78 μ V | |
| | (1 to 40) Hz | 61.4 μ V/V + 81.4 μ V | |
| AC Voltage Measure (0.1 to 1.2) V | 40 Hz to 1 kHz | 49 μ V/V + 49.2 μ V | Agilent 3458A |
| | (1 to 20) kHz | 128 μ V/V + 44.1 μ V | |
| | (20 to 50) kHz | 277 μ V/V + 47.1 μ V | |
| | (50 to 100) kHz | 745 μ V/V + 61.5 μ V | |
| | (100 to 300) kHz | 2.9 mV/V + 223 μ V | |
| | (0.3 to 1) MHz | 9.4 mV/V + 518 μ V | |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

Electrical – Voltage

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|--------------------------------------|------------------|---|-------------------|
| AC Voltage Measure (1 to 12) V | (1 to 40) Hz | 61.4 μ V/V + 814 μ V | Agilent 3458A |
| | 40 Hz to 1 kHz | 64.5 μ V/V + 445 μ V | |
| | (1 to 20) kHz | 127 μ V/V + 442 μ V | |
| | (20 to 50) kHz | 277 μ V/V + 471 μ V | |
| | (50 to 100) kHz | 745 μ V/V + 615 μ V | |
| | (100 to 300) kHz | 2.9 mV/V + 2.3 mV | |
| | (0.3 to 1) MHz | 9.4 mV/V + 5.2 mV | |
| AC Voltage Measure (10 to 120) V | (1 to 40) Hz | 175 μ V/V + 8.8 mV | Agilent 3458A |
| | 40 Hz to 1 kHz | 184 μ V/V + 4.4 mV | |
| | (1 to 20) kHz | 184 μ V/V + 4.4 mV | |
| | (20 to 50) kHz | 324 μ V/V + 4.8 mV | |
| | (50 to 100) kHz | 1.2 mV/V + 6.8 mV | |
| | (100 to 300) kHz | 3.8 mV/V + 29 mV | |
| AC Voltage Measure (100 to 700) V | (1 to 40) Hz | 465 μ V/V + 74.5 mV | Agilent 3458A |
| | 40 Hz to 1 kHz | 444 μ V/V + 33.2 mV | |
| | (1 to 20) kHz | 672 μ V/V + 33.8 mV | |
| | (20 to 50) kHz | 1.4 mV/V + 44.8 mV | |
| | (50 to 100) kHz | 3.5 mV/V + 44.8 mV | |
| 700V to 5 kV | (45 to 500) Hz | 15.1 mV/V + 6.2 V | Fluke 80k-6 / 87V |

| PARAMETER EQUIPMENT | RANGE | CALIBRATION & MEASUREMENT CAPABILITIES EXPRESSED AS UNCERTAINTY (\pm) | REFERENCE STANDARD OR EQUIPMENT |
|--|---|---|---------------------------------------|
| Electrical Simulation of Thermocouple | | | |
| Indicators | | | |
| Type T | -250 to -150 °C -150 to 0 °C 0 to 120 °C 120 to 400 °C | 0.63 °C 0.24 °C 0.16 °C 0.12 °C | |
| Type C | 0 to 15 °C 150 to 650 °C 650 to 1 000 °C 1 000 to 1 800 °C 1 800 to 2 316 °C | 0.30 °C 0.26 °C 0.31 °C 0.50 °C 0.84 °C | |
| Type E | -250 to -100 °C -100 to -25 °C -25 to 350 °C 350 to 650 °C 650 to 1 000 °C | 0.50 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C | Fluke 5500A |
| Type J | -210 to -100 °C -100 to -30 °C -30 to 150 °C 150 to 760 °C 760 to 1 200 °C | 0.27 °C 0.16 °C 0.14 °C 0.17 °C 0.23 °C | |
| Type K | 200 to -100 °C -100 to -25 °C -25 to 120 °C 120 to 1 000 °C 1 000 to 1 372 °C | 0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.40 °C | |
| Type L | 200 to -100 °C -100 to 800 °C 800 to 900 °C | 0.37 °C 0.26 °C 0.17 °C | |

Electrical - Current

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---------------------------------|--------------------|---|-----------------------------|
| DC Current - Source | (0 to 200) μ A | 111 μ A/A + 2.4 nA | Wavetek 4808 Opt 10 / 40 |
| | (0.1 to 2.0) mA | 44.4 μ A/A + 12.4 nA | |
| | (1.0 to 20) mA | 44 μ A/A + 130 nA | |
| | (10 to 200) mA | 44.6 μ A/A + 1.3 μ A | |
| | (0.1 to 2) A | 110 μ A/A + 25.2 μ A | |
| | (1 to 20) A | 711 μ A/A + 11 mA | |
| DC Current - Measure | (0 to 120) μ A | 15.4 μ A/A + 2.0 nA | Agilent 3458A |
| | (0.1 to 1.2) mA | 15.1 μ A/A + 15.3 nA | |
| | (1.0 to 12) mA | 14.4 μ A/A + 161 nA | |
| | (10 to 120) mA | 29.1 μ A/A + 2 μ A | |
| | (0.1 to 1.05) A | 98.1 μ A/A + 24.3 μ A | |
| | (1 to 20) A | 711 μ A/A + 11 mA | |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

Electrical – Current

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---------------------------------|-----------------|---|-----------------------------|
| AC Current - Source | 10 Hz to 1 kHz | 210 μ A/A + 18.7 nA | Wavetek 4808 Opt 20 / 40 |
| | (1 to 5) kHz | 336 μ A/A + 26.4 nA | |
| | 10 Hz to 1 kHz | 171 μ A/A + 220 nA | |
| | (1 to 5) kHz | 288 μ A/A + 215 nA | |
| | (1 to 20) mA | 114 μ A/A + 1.3 μ A | |
| | (1 to 5) kHz | 203 μ A/A + 1.2 μ A | |
| | (10 to 200) mA | 128 μ A/A + 16.9 μ A | |
| | (1 to 5) kHz | 223 μ A/A + 16.8 μ A | |
| | (0.1 to 2) A | 358 μ A/A + 163 μ A | |
| | (1 to 5) kHz | 526 μ A/A + 197 μ A | |
| AC Current - Source | (10 to 40) Hz | 1.6 mA/A + 1.8 mA | Wavetek 9100 |
| | 40 Hz to 1 kHz | 1.6 mA/A + 1.7 mA | |
| | (1 to 5) kHz | 5.5 mA/A + 27.1 mA | |
| AC Current - Measure | (10 to 20) Hz | 3.8 mA/A + 50.6 nA | Agilent 3458A |
| | (20 to 45) Hz | 1.5 mA/A + 45.6 nA | |
| | (45 to 100) Hz | 576 μ A/A + 43.8 nA | |
| | 100 Hz to 1 kHz | 565 μ A/A + 43.1 nA | |
| AC Current - Measure | (10 to 20) Hz | 3.8 mA/A + 390 nA | |
| | (20 to 45) Hz | 1.4 mA/A + 355 nA | |
| | (45 to 100) Hz | 521 μ A/A + 348 nA | |
| | 100 Hz to 5 kHz | 271 μ A/A + 301 nA | |
| | (5 to 20) kHz | 519 μ A/A + 348 nA | |
| | (20 to 50) kHz | 3.8 mA/A + 665 nA | |
| AC Current - Measure | (10 to 20) Hz | 3.8 mA/A + 3.9 μ A | |
| | (20 to 45) Hz | 1.4 mA/A + 3.6 μ A | |
| | (45 to 100) Hz | 540 μ A/A + 3.5 μ A | |
| | 100 Hz to 5 kHz | 277 μ A/A + 3 μ A | |
| | (5 to 20) kHz | 538 μ A/A + 3.5 μ A | |
| | (20 to 50) kHz | 3.8 mA/A + 6.7 μ A | |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

Electrical – Current

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---|-----------------|---|----------------|
| AC Current - Measure (10 to 120) mA | (10 to 20) Hz | 3.8 mA/A + 39 µA | Agilent 3458A |
| | (20 to 45) Hz | 1.4 mA/A + 35.4 µA | |
| | (45 to 100) Hz | 540 µA/A + 34.6 µA | |
| | 100 Hz to 5 kHz | 277 µA/A + 30 µA | |
| | (5 to 20) kHz | 538 µA/A + 34.6 µA | |
| | (20 to 50) kHz | 3.8 mA/A + 66.4 µA | |
| AC Current - Measure (0.1 to 1.05) A | (10 to 20) Hz | 3.8 mA/A + 362 µA | 3458A / EL7520 |
| | (20 to 45) Hz | 1.5 mA/A + 336 µA | |
| | (45 to 100) Hz | 717 µA/A + 329 µA | |
| | 100 Hz to 5 kHz | 932 µA/A + 282 µA | |
| | (5 to 20) kHz | 2.8 mA/A + 349 µA | |
| AC Current - Measure (1 to 20) A | (10 to 40) Hz | 175 µA/A + 219 µA | 3458A / EL7520 |
| | 40 Hz to 1 kHz | 152 µA/A + 311 µA | |
| | (1 to 5) kHz | 159 µA/A + 477 µA | |

Electrical – Resistance

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---|--------|---|----------------------------|
| Resistance - Source Fixed Points | 10 Ω | 0.31 mΩ | Wavetek 4808 Opt 50 |
| | 100 Ω | 1.2 mΩ | |
| | 1 kΩ | 11.8 mΩ | |
| | 10 kΩ | 118 mΩ | |
| | 100 kΩ | 1.3 Ω | |
| | 1 MΩ | 30.9 Ω | |
| | 10 MΩ | 612 Ω | |
| | 100 MΩ | 10.4 kΩ | |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

Electrical – Resistance

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---------------------------------|---------------------|---|---------------|
| Resistance - Source Adjustable | (0.1 to 10.999) Ω | 139 μΩ/ Ω + 7 mΩ | Fluke 5500A |
| | (11 to 32.999) Ω | 138 μΩ/ Ω + 11.7 mΩ | |
| | (33 to 109.999) Ω | 104 μΩ/ Ω + 11.7 mΩ | |
| | (110 to 329.999) Ω | 107 μΩ/ Ω + 11.3 mΩ | |
| | 330Ω to 1.0999 kΩ | 36.2 μΩ/ Ω + 31.4 mΩ | |
| | (1.1 to 3.2999) kΩ | 109 μΩ/ Ω + 64.6 mΩ | |
| | (3.3 to 10.9999) kΩ | 104 μΩ/ Ω + 697 mΩ | |
| | (11 to 32.9999) kΩ | 108 μΩ/ Ω + 651 mΩ | |
| | (33 to 109.999) kΩ | 126 μΩ/ Ω + 7.1 Ω | |
| | (110 to 329.999) kΩ | 147 μΩ/ Ω + 6.1 Ω | |
| | 330kΩ to 1.0999M Ω | 172 μΩ/ Ω + 65.1 Ω | |
| | (1.1 to 3.2999) MΩ | 218 μΩ/ Ω + 57.6 Ω | |
| | (3.3 to 10.9999) MΩ | 693 μΩ/ Ω + 662 Ω | |
| | (11 to 32.9999) MΩ | 1.3 mΩ/ Ω + 4.3 kΩ | |
| | (33 to 109.999) MΩ | 5.6 mΩ/ Ω + 6.6 kΩ | |
| Resistance Measure | (1 to 12) Ω | 12.8 μΩ/ Ω + 180 μΩ | Agilent 3458A |
| | (10 to 120) Ω | 9.6 μΩ/ Ω + 1.6 mΩ | |
| | 100 Ω to 1.2 kΩ | 7.2 μΩ/ Ω + 6.6 mΩ | |
| | (1 to 12) kΩ | 7.1 μΩ/ Ω + 65.9 mΩ | |
| | (10 to 120) kΩ | 7.2 μΩ/ Ω + 647 mΩ | |
| | 100 kΩ to 1.2 MΩ | 11.7 μΩ/ Ω + 10 Ω | |
| | (1 to 12) MΩ | 40 μΩ/ Ω + 299 Ω | |
| | (10 to 120) MΩ | 501 μΩ/ Ω + 5.7 kΩ | |

Time and Frequency - Frequency / Period

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---------------------------------|--------------------|---|-------------------|
| Frequency Source | 10 MHz Fixed | 2.8 parts in 10^{11} | GPSDO Z3805A |
| | (0.1 to 30) Hz | 3.7 parts in 10^5 | |
| | 30 Hz to 100 kHz | 1.3 parts in 10^7 | Agilent 3325A |
| | (0.1 to 20) MHz | 5.8 parts in 10^9 | |
| | 20 MHz to 2.05 GHz | 1.2 parts in 10^9 | Agilent 8644A |
| | (2 to 26) GHz | 6.3 parts in 10^{10} | Agilent |
| | (26 to 50) GHz | 2.3 parts in 10^{10} | 83650B |
| Frequency Measure | (0.1 to 30) Hz | 3.6 parts in 10^5 | Agilent 53132A |
| | 30 Hz to 200 kHz | 5.4 parts in 10^8 | |
| | (0.2 to 100) MHz | 1.2 parts in 10^9 | |
| | 100 MHz to 3 GHz | 2.4 parts in 10^{10} | |
| | (3 to 26) GHz | 6.4 parts in 10^{10} | Agilent 5351B |
| | (26 to 50) GHz | 7.5 parts in 10^{10} | Agilent 8563E |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

| | | | |
|---|---|--|--------------------------|
| Oscilloscope | | | |
| Amplitude DC Signal into 50 Ω load into 1M Ω load | -2.2 to 2.2 V -33 to 33 V | 2.0 mV/V + 100 µV 2.0 mV/V + 100 µV | |
| Amplitude Squarewave 50 Ω load 1MΩ load | 1.8 mV to 2.2 V p-p 10 Hz to 10 kHz 1.8 mV to 5 V p-p 10 Hz to 10 kHz 95 To 105 V p-p 100 to 1000 Hz | 2.1 mV/V + 100 µV .1 mV/V + 100 µV 3.9 mV/V + 100 µV | Fluke 5500A opt SC300 |
| Rise Time | 305 pS | 65 pS | |
| Leveled Sine Wave Relative to 50 kHz 5 mV to 5.5 V | 50 kHz to 100 MHz 100 to 300 MHz | 35 mV/V + 300 µV 40 mV/V + 300 µV | |
| Time Marker | 2 nS to 5 S 5 - 2 - 1 | 0.78 mSec / Sec | Fluke 5500A opt SC300 |
| Wave Generator - Source Amplitude 10 Hz to 10 kHz square, sine, triangle into 1 MΩ square, sine, triangle into 50 Ω | 1.8 mV to 55 V p-p 1.8 mV to 2.2 V p-p | 30 mV/V + 100 µV 30 mV/V + 100 µV | |

Electrical-RF / Microwave

| PARAMETER EQUIPMENT | RANGE | CALIBRATION & MEASUREMENT CAPABILITIES EXPRESSED AS UNCERTAINTY (\pm) | REFERENCE STANDARD OR EQUIPMENT |
|--|---|---|---|
| Generators | | | |
| Tune RF Level – Measure, 0 to -120 dB | | | |
| 0 to 10 dB | 2 MHz to 26.5 GHz | 0.22 dB | |
| 0 to 20 dB | | 0.23 dB | |
| 0 to 30 dB | | 0.23 dB | |
| 0 to 40 dB | | 0.23 dB | |
| 0 to 50 dB | | 0.24 dB | Agilent 8902A |
| 0 to 60 dB | | 0.24 dB | W/ 11722A & |
| 0 to 70 dB | | 0.25 dB | Microwave Mixer |
| 0 to 80 dB | | 0.25 dB | |
| 0 to 90 dB | | 0.25 dB | |
| 0 to 100 dB | | 0.26 dB | |
| 0 to 110 dB | | 0.27 dB | |
| 0 to 120 dB | | 0.29 dB | |
| Modulation – Measure | | | |
| Phase Rates 200Hz to 10 kHz | 0.15 to 10 MHz 10 to 1300 MHz 1.3 GHz to 26.5 GHz | 5.0% 4.0% 5.0% | |
| Frequency Rates 50Hz to 10 kHz | 0.25 to 10 MHz 10 to 1300 MHz 1.3 GHz to 26.5 GHz | 3.0% 2.0% 3.0% | Agilent 8902A W/ 11722A Microwave Mixer |
| Amplitude Rates 50Hz to 10 kHz | 0.15 to 10 MHz 10 to 1300 MHz 1.3 GHz to 26.5 GHz | 3.0% 2.0% 3.0% | |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

| PARAMETER EQUIPMENT | RANGE | CALIBRATION & MEASUREMENT CAPABILITIES EXPRESSED AS UNCERTAINTY (\pm) | REFERENCE STANDARD OR EQUIPMENT |
|--|--|---|---|
| Generators Continued | | | |
| Spectrum Analysis | 100 Hz to 50 GHz | 1.5 dB | Agilent 8563E Agilent 11970 Mixer |
| Total Harmonic Distortion | 20 Hz to 20 kHz 20 kHz to 100 kHz | 1.2 dB 2.2dB | Agilent 8903B |
| Power Accuracy Frequency Response/Power Accuracy 1 uW - 100 mW | 1 mW 50 MHz 0.1 to 1000 MHz 1 to 12 GHz 12 to 18 GHz 18 to 26.5GHz 26.5 to 50 GHz | 1.3% 4.3% 4.3% 4.5% 4.8% 5.1% | 478A H76 8482A 8481A 8481A 8485A 8487A |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

| PARAMETER EQUIPMENT | RANGE | CALIBRATION & MEASUREMENT CAPABILITIES EXPRESSED AS UNCERTAINTY (\pm) | REFERENCE STANDARD OR EQUIPMENT |
|--|--|---|---------------------------------------|
| Leveled Sine Wave Relative to 50 kHz 5 mV to 5.5 V | 50 kHz to 100 MHz 100 to 300 MHz | 35 mV/V + 300 μ V 40 mV/V + 300 μ V | |
| Time Marker | 2 nS to 5 S 5 - 2 - 1 | 0.78 mSec / Sec | Fluke 5500A opt SC300 |
| Wave Generator - Source Amplitude 10 Hz to 10 kHz square, sine, triangle into 1 M Ω square, sine, triangle into 50 Ω | 1.8 mV to 55 V p-p 1.8 mV to 2.2 V p-p | 30 mV/V + 100 μ V 30 mV/V + 100 μ V | |
| Pulse Measurement | | | |
| Transition Time | 10 to 90 % | 17.5 pS | Agilent 54120T |
| Pulse Width | \leq 10 ns $>$ 10 ns $>$ 50 μ s $>$ 500 μ s | 5.0% + 10 pS 1.56% 0.20% 0.02% | Agilent 53131A |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

| PARAMETER EQUIPMENT | RANGE | CALIBRATION & MEASUREMENT CAPABILITIES EXPRESSED AS UNCERTAINTY (\pm) | REFERENCE STANDARD OR EQUIPMENT |
|---|---|---|---|
| Spectrum Analyzers | | | |
| Frequency Readout Accuracy | 100 Hz to 50GHz | 2 Count of LSD of rdg | Agilent 8663A / 83650B |
| Amplitude Accuracy and Frequency Response | | | |
| 50Ω | 1 mHz to 200 Hz 200 Hz to 100 kHz 100 kHz to 12 GHz 12 to 18 GHz 18 to 26.5 GHz 26.5 to 50 GHz | 0.15 dB 0.12 dB 0.19 dB 0.21 dB 0.22 dB 0.24 dB | Agilent E4419A 848XX Sensor 3325A, 3335A |
| 75Ω | 200 Hz to 1 kHz 1 kHz to 25 MHz 25 to 80 MHz | 0.17 dB 0.09 dB 0.17 dB | Agilent 3335A |
| Scale Fidelity | | | |
| 50 Ω ref to 100 kHz | 0 to 18 dB 20 to 58 dB 60 to 98 dB | 0.06 dB 0.10 dB 0.21 dB | |
| 75 Ω ref to 100 kHz | 0 to 18 dB 20 to 58 dB 60 to 98 dB | 0.16 dB 0.26 dB 0.51 dB | Agilent 3335A |
| Span | 1 Hz to 50 GHz | 2 Count of LSD of rdg | Agilent 83650B Agilent 8663A |
| Resolution Bandwidth | 1 Hz to 10 MHz | 2 Count of LSD of rdg | Agilent 3335A |

TRIDENT SYSTEMS & ENGINEERING

APPENDIX
July 2012

| PARAMETER EQUIPMENT | RANGE | CALIBRATION & MEASUREMENT CAPABILITIES EXPRESSED AS UNCERTAINTY (\pm) | REFERENCE STANDARD OR EQUIPMENT |
|--|---|---|---------------------------------------|
| Scattering Parameters (Worst Case) Reflection: S11 and S22 Reflection Coefficient 0 to 1 | 0.3 to 45 MHz .045 to 2 GHz 2 to 8 GHz 8 to 20 GHz 20 to 26.5 GHz | 0.05 lin 6.5° 0.02 lin 2.7° 0.03 lin 4.3° 0.04 lin 4.5° 0.04 lin 4.8° | Agilent 8753C Agilent 8510C |
| Scattering Parameters – Transmission:S21 and S12 0 to -40 | 0.3 to 45 MHz .045 to 2 GHz 2 to 8 GHz 8 to 20 GHz 20 to 26.5 GHz | 0.2 dB 2.5° 0.07 dB 0.5° 0.07 dB 1.3° 0.11 dB 2.8° 0.23 dB 4.2° | Agilent 8753C Agilent 8510C |

Length - Artifacts and Standards 1D

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---------------------------------|--------------------|---|------------------|
| Plugs Cylindrical ¹ | (0.011 to 0.75) in | 35 μin | Super micrometer |

Length - Hand Tools and Precision Gages 1D

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---------------------------------|-----------------|---|-------------|
| Calipers ¹ | (0.05 to 60) in | (330 + 3L) μin | Gage Blocks |
| Micrometers ¹ | (0.05 to 12) in | (29 + 5L) μin | Gage Blocks |
| Height Gages ¹ | (0.05 to 60) in | (330 + 3L) μin | Gage Blocks |
| Indicators ¹ | (0.05 to 2) in | (14 + 23L) μin | Gage Blocks |

| PARAMETER EQUIPMENT | RANGE | CALIBRATION & MEASUREMENT CAPABILITIES EXPRESSED AS UNCERTAINTY (\pm) ² | REFERENCE STANDARD OR EQUIPMENT |
|---------------------|--------------------------------|--|--------------------------------------|
| Gage Blocks | (0 to 0.1) in (0.1 to 4) in | 4.0 μin (3.4 + 3.0L) μin | Gage Block Comparator Gage Blocks |
| Micrometers Depth | (0 to 12) in | (65 + 1.5L) μin | Gage Blocks |
| Micrometers Inside | (3 to 12) in | (58 + 1.7L) μin | Gage Blocks |

Mass - Pressure/Low Vacuum

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|---|------------------|---|---------------------|
| Pressure Gages and Transducers ¹ | (0 to 300) psi | 0.17 psi + 0.0007 psi/psi | Pressure Calibrator |
| | (5 to 5 000) psi | 0.02 psi + 0.00007 psi/psi | Dead weight Tester |
| Vacuum Gages and Transducers ¹ | (0 to 29) in Hg | 0.38 in Hg | Pressure Calibrator |

Mass - Scale and Balances

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|----------------------------------|---------------|---|------------------|
| Analytical Balances ¹ | (0 to 200) g | 0.62 mg | Class S Masses |
| Weigh Scales ¹ | (0 to 100) lb | 0.32 g + 0.08 mg/g | Class 6/7 Masses |

Mass - Torque

| Calibration Parameter/Equipment | Range | Calibration and Measurement Capability(+/-) | Remarks |
|--|-------------------|---|-------------------------|
| Torque Watches & Screwdrivers ¹ | (2 to 215) ozf·in | 0.07 ozf·in + 0.009 ozf·in/ozf·in | Torque Watch Calibrator |
| Torque Wrenches ¹ | (1 to 100) lbf·ft | 0.1 lbf·in + 0.01 lbf·in/ lbf·in | Torque Calibrator |

Notes

1 ISO IEC 17025 Accredited calibrations available

2 L = length in inches,