

TSE



Certificate of Calibration



Trident Systems Engineering
2646 Palma Dr. Ste. 130
Ventura, Ca. 93003
805 - 830 - 8596

Certificate 2190091

Based on a Recommended/ agreed on Cal interval of 12 Months

Your Company Name

Test Date of Calibration 16 Jan 2019 The Recall Date is 16 Jan 2020

Your Address

Your City State Zip

PO Number Your PO Number

Manufacturer Fluke

Cal Location In Laboratory

Model 73 III

Procedure 33K8-4-14-1-122015

Description Digital Multimeter

Technician 10 Quality Assurance

Serial Number Serial Number

Temperature 22 °C

Asset Number Your Asset Number

Humidity: 45 %



Received Condition

Pass + - The measured values of the equipment were observed in specification at the points tested. Additionally, the expanded measurement uncertainty intervals about the measured values could have been out of specification with a PFA of <6.0%

Returned Condition

As Received


Cleaned and Calibrated to Manufacturer's Specifications in accordance with the procedure listed above

See Attached Data

This certifies that the above listed product has been calibrated with a quality system accredited to ISO / IEC 17025 : 2017
This Calibrations report is composed of a certificate of calibration, performance test results and/or certificate appendice
Each Report section may be numbered separately. Estimated Measurement Uncertainty is shown on the calibration Report
Traceability is through NIST or another National Metrology Institute to the International System of Units (SI units).
Some measurements are traceable to natural physical constants, Consensus standards or ratio type measurements
supporting documentation relative to traceability is on file and available for examination upon request
Although fully traceable measurements marked with an asterisk(*) are outside our Scope of Accreditation
This Certificate shall not be reproduced, except in full, without written approval by TSE

Measurement results are reported as:

- Pass -The measured values of the equipment were observed in specification at the points tested. Additionally, the expanded measurement uncertainty intervals about the measured values were in specification.
- Pass + -The measured values of the equipment were observed in specification at the points tested. Additionally, the expanded measurement uncertainty intervals about the measured values could have been out of specification with a PFA of <6.0%
- Fail + -The measured values of the equipment were observed out of specification at the points tested. Additionally, the expanded measurement uncertainty intervals about the measured values could have been in specification with a PFR of <10.0%
- Fail -One or more measured values of the equipment were observed out of specification at the points tested. Additionally, the expanded measurement uncertainty intervals about one or more measured values were entirely outside the specification



$TUR \geq 1.5 : 1$

Zero Guardbanding Employed

Pass < 0.15 % Probability of False Accept (PFA)

Pass + < 6.0 % Probability of False Accept (PFA)

Fail + < 10.0 % Probability of False Reject (PFR)

Fail < 0.15 % Probability of False Reject (PFR)

Standards used in this Calibration

Asset Number	Model Number	Description	Recall Date	Trace Number
TR204	4808	Multifunction Calibrator	10 Dec 2019	1002212237
TR112	5500A	Multi-Product Calibrator	29 Dec 2019	2183022

Certificate Number 2190091DATE 19 Jan, 2019

2646 Palma Dr. #130

Ventura, Ca. 93003

Website TSECAL.com

Phone 805-830-8596

Fax 805-642-2259

TRIDENT SYSTEM & ENGINEERING

Customer Your Company NameTEMP. °C 22R.H. % 45Manufacturer FlukeProcedure 33K8-4-14-1-122015As Received XModel NO. 73 III

Post Adjustment and /or Repair _____

Description Digital MultimeterSerial NO. Serial NumberAsset NO. Your Asset Number

VERIFICATION TEST PERFORMED:

TEST	Nominal		Minimum	Measured Reading	Maximum	Manufacturer Specification	EMU ±	Acceptance Criteria	
PASS/ FAIL									
4.1 DC Volts Calibration									
	300	mV	299.0	299.0	301.0	1.0	1.0E-01	PASS	+
	3	V	2.990	3.010	3.010	0.010	1.0E-03	PASS	+
	10	V	9.96	9.96	10.04	0.04	1.0E-02	PASS	+
	-10	V	-10.04	-9.96	-9.96	0.04	1.0E-02	PASS	+
	20	V	19.93	19.93	20.07	0.07	1.0E-02	PASS	+
	-20	V	-20.07	-19.93	-19.93	0.07	1.0E-02	PASS	+
	30	V	29.90	30.10	30.10	0.10	1.0E-02	PASS	+
	-30	V	-30.10	-30.10	-29.90	0.10	1.0E-02	PASS	+
	300	V	299.0	299.0	301.0	1.0	1.0E-01	PASS	+
	550	V	547	553	553	3	1.0E+00	PASS	+
4.2 AC Volts Calibration									
45 Hz	3	V	2.938	2.938	3.062	0.062	3.0E-03	PASS	+
500 Hz	3	V	2.938	3.062	3.062	0.062	3.0E-03	PASS	+
45 Hz	30	V	29.38	29.38	30.62	0.62	3.0E-02	PASS	+
1 kHz	30	V	29.38	30.62	30.62	0.62	3.0E-02	PASS	+
45 Hz	300	V	293.8	293.8	306.2	6.2	3.0E-01	PASS	+
1 kHz	300	V	293.8	306.2	306.2	6.2	3.0E-01	PASS	+
45 Hz	550	V	537	537	563	13	1.0E+00	PASS	+
1 kHz	550	V	537	563	563	13	1.0E+00	PASS	+
4.3 DC Current Calibration									
	30	mA	29.53	29.53	30.47	0.47	1.0E-02	PASS	+
	300	mA	293.8	306.2	306.2	6.2	1.0E-01	PASS	+
	9.5	A	9.34	9.34	9.66	0.16	1.0E-02	PASS	+
4.4 AC Current Calibration									
45 Hz	30	mA	29.23	30.77	30.77	0.77	6.0E-02	PASS	+
1 kHz	30	mA	29.23	29.23	30.77	0.77	6.0E-02	PASS	+
45 Hz	300	mA	292.30	307.7	307.70	7.70	6.0E-01	PASS	+
1 kHz	300	mA	292.30	292.3	307.70	7.70	6.0E-01	PASS	+
45 Hz	9.5	A	9.25	9.25	9.75	0.25	2.0E-02	PASS	+
1 kHz	9.5	A	9.25	9.75	9.75	0.25	2.0E-02	PASS	+
4.5 Resistance Calibration									
	190	Ω	188.8	188.8	191.2	1.2	4.0E-01	PASS	+
	1900	Ω	1889	1911	1911	11	4.0E+00	PASS	+
	19	kΩ	18.89	18.89	19.11	0.11	4.0E-02	PASS	+
	190	kΩ	188.9	191.1	191.1	1.1	4.0E-01	PASS	+
	1.9	MΩ	1.889	1.889	1.911	0.011	4.0E-03	PASS	+
	19	MΩ	18.61	19.39	19.39	0.39	6.0E-02	PASS	+